



# CERTIFICATION

**AOAC<sup>®</sup> Performance Tested<sup>SM</sup>**

Certificate No.

**010401**

The AOAC Research Institute hereby certifies the test kit known as:

**Nissui Compact Dry Total Count**

manufactured by

**NISSUI Pharmaceutical Co., Ltd.**

**3-24-6, Ueno**

**Taito-ku, Tokyo**

**Japan 110-8736**

This method has been evaluated in the AOAC<sup>®</sup> *Performance Tested Methods*<sup>SM</sup> Program and found to perform as stated by the manufacturer contingent to the comments contained in the manuscript. This certificate means that an AOAC<sup>®</sup> Certification Mark License Agreement has been executed which authorizes the manufacturer to display the AOAC *Performance Tested*<sup>SM</sup> certification mark along with the statement - "THIS METHOD'S PERFORMANCE WAS REVIEWED BY AOAC RESEARCH INSTITUTE AND WAS FOUND TO PERFORM TO THE MANUFACTURER'S SPECIFICATIONS" - on the above mentioned method for a period of one calendar year from the date of this certificate (December 26, 2019 – December 31, 2020). Renewal may be granted at the end of one year under the rules stated in the licensing agreement.

*Scott Coates*

\_\_\_\_\_  
Scott Coates, Senior Director  
Signature for AOAC Research Institute

December 26, 2019

\_\_\_\_\_  
Date

<b>METHOD AUTHORS</b> ORIGINAL VALIDATION: Hidemasa Kodaka, Shingo Mizuochi, Hajime Teramura, and Tadanobu Nirazuka MODIFICATION APRIL 2015: Shingo Mizuochi	<b>SUBMITTING COMPANY</b> NISSUI Pharmaceutical Co., Ltd. 3-24-6, Ueno Taito-ku, Tokyo Japan 110-8736
--	---

<b>KIT NAME(S)</b> Nissui Compact Dry Total Count	<b>CATALOG NUMBERS</b> Nissui 06740 (40 plates), 06741 (240 plates)
--	--

<b>INDEPENDENT LABORATORY</b> ORIGINAL VALIDATION National Public Health Service for Wales Llandough Hospital Penlan Rd, Penarth, United Kingdom	<b>MODIFICATION APRIL 2015</b> Campden BRI Station Road Chipping Campden Gloucestershire, GL55 6LD, UK	<b>EXPERT REVIEWERS</b> ORIGINAL VALIDATION: AOAC RI PTM reviewers Yi Chen <sup>1</sup> , Yvonne Salfinger <sup>2</sup> , Maria Christina Fernandez <sup>3</sup> <sup>1</sup> US FDA CFSAN, College Park, MD, USA <sup>2</sup> Consultant, Denver, CO, USA <sup>3</sup> University of Buenos Aires, Buenos Aires, ARGENTINA
--	--	--

<b>APPLICABILITY OF METHOD</b> Target Organism – Aerobic microbial cells  Matrices – raw meat (50 g) (raw ground pork, raw pork, raw lamb, raw veal, raw ground beef) MODIFICATION April 2015 - cooked chicken (10 g), lettuce (10 g), milk powder (10 g), frozen fish (10 g), pasteurized milk (1 mL)  Performance claims - Performance is equivalent to the reference method.	<b>REFERENCE METHODS</b> <i>Official Methods of Analysis</i> 19 <sup>th</sup> Ed., AOAC INTERNATIONAL, Rockville, MD, Method 966.23, <i>Microbiological Methods</i> , <a href="http://www.eoma.aoac.org">http://www.eoma.aoac.org</a> (2) ISO 4833:2003, <i>Microbiology of food and animal feeding stuffs – Horizontal Method for the enumeration of microorganisms – Colony-count technique at 30°C.</i> (3)
---	--

<b>ORIGINAL CERTIFICATION DATE</b> February 2004	<b>CERTIFICATION RENEWAL RECORD</b> Renewed annually through December 2020
---	---

<b>METHOD MODIFICATION RECORD</b> 1. April 2015 Level 2  2. February 2019 Level 2	<b>SUMMARY OF MODIFICATION</b> 1. Matrix extension to include cooked chicken, pre-washed bagged shredded iceberg lettuce, frozen cod fillets, instant non-fat dry milk, pasteurized milk (2%). 2. Shelf life increased to 24 months and corporate address change.
--	---

Under this AOAC® <i>Performance Tested</i> <sup>SM</sup> License Number, 010404 this method is distributed by: 1. Hardy Diagnostics	Under this AOAC® <i>Performance Tested</i> <sup>SM</sup> License Number, 010404 this method is distributed as: 1. CompactDry™ TC
--	---

#### PRINCIPLE OF THE METHOD (1)

Compact Dry TC qualifies as a rapid method kit for determining aerobic colony counts in foods. The plates are presterilized and contain culture medium and a cold-soluble gelling agent. The medium is rehydrated by inoculating 1 mL diluted sample into the center of the self-diffusible medium and allowing the solution to diffuse by capillary action. The plates can then be incubated and the colonies counted without any additional steps. The Compact Dry TC method was validated with 5 different raw meats. The performance tests were conducted at 35°C and 30°C. In all required performance studies, no apparent differences were observed between the Compact Dry TC method and the Standard Pour Plate method (AOAC Official Method 966.23) for the detection level of aerobic microorganisms. For the accuracy claim ( $n = 60$ ), a correlation factor of  $r^2_{35} = 0.9977$  (35°C) and  $r^2_{30} = 0.9932$  (30°C) could be assigned, as stated in the application for "Performance Tested Method<sup>SM</sup>". Quality consistency and storage robustness studies, showed no significant variations in plate count results with different production lots or plates of diverse storage age.

**DISCUSSION OF THE VALIDATION STUDY (1)**

The retested results, as instructed by AOAC RI with exactly the same sample preparation procedure as AOAC Official Method **966.23** showed equal results on both the Compact Dry TC Method and AOAC Official Method **966.23**. As evidenced by attached data of 5 replicates for 3 levels (4 levels for raw veal) of 5 kinds of raw meat ( $n = 80$ ) through internal and independent validation study,  $r^2$  as correlation coefficient for both methods was 0.9965. This means that the retested results are the same as the original, and therefore discussion on the original report must be valid.

Regarding the measured value, for only the lowest level, the value between Compact Dry TC and AOAC Official Method **966.23** differed from each other more than the range of the standard deviation, and by one-way ANOVA ( $P < 0.05$ ) in independent validation studies. This phenomenon could not be observed with the other bacterial load levels and with the other 4 matrixes (raw ground pork, raw pork, raw lamb, and raw veal) by internal validation study. For all remaining levels and samples by internal validation study, the standard deviation of analogous samples delivered comparable values for the Compact Dry TC method and the AOAC Official Method **966.23**. Consequently, the results from the Compact Dry TC method for the repeatability study correspond with the results from pour plate methods.

Using an incubation temperature of 30°C instead of the recommended 35°C, which is the more common incubation temperature for total aerobic count estimations in the food regulations of the European Community, the results of the repeatability, accuracy, lot consistency, and lot stability obtained from the different matrixes were similar to those at 35°C. Using yogurt in the repeatability study, the 30°C incubation series exhibited major differences. The counts of the dilutions 10<sup>4</sup> and 10<sup>3</sup> measured with Compact

Dry TC were significantly lower than those with the pour plate method. Dilution 2.5 x 10<sup>2</sup> and 10<sup>2</sup> did not give any count with the Compact Dry TC technique.

Because those results with yogurt could not be observed in the accuracy, lot consistency, and lot stability study at 30°C incubation, the results of the repeatability study might be attributed to internal errors.

The big advantages of the Compact Dry TC system are the reduced hands-on time and economical usage, as confirmed by the independent laboratory. In terms of plate preparation, inoculation, and reading the result, the Compact Dry TC system was easier and quicker than the conventional pour plate technique. Reading the plates was faster with the Compact Dry TC system, with the TTC indicator speeding up counting. It was observed that food particles, when present, did not appear to absorb the indicator. Food particles in the pour plate made reading plates and counting colonies relatively more difficult. For the Compact Dry TC system, less training is required than for the pour plate technique. Instructions on the use of the Compact Dry TC are clear and unambiguous. The Compact Dry TC system would also bring advantages in reduced storage space, waste disposal, and required incubator space. The long shelf life of the product also has benefits compared to ready-prepared agar, which has a limited shelf life and therefore requires more logistical planning.

**ORIGINAL VALIDATION DATA**

**Table 1. AOAC method comparison (raw ground pork) (1)**

Level	Spike, CFU/g	Replicate	Compact Dry TC aerobic bacteria		AOAC <b>966.23</b> aerobic bacteria	
			CFU/g	Log <sub>10</sub> CFU/g	CFU/g	log <sub>10</sub> CFU/g
About 10 <sup>2</sup> CFU/g Lot 1	0 <sup>a</sup>	1	525	2.72	500	2.70
		2	680	2.83	670	2.83
		3	430	2.63	415	2.62
		4	635	2.80	635	2.80
		5	365	2.56	355	2.55
		Mean	527	2.71	515	2.70
		s <sub>r</sub>	133	0.11	136	0.12
		RSD <sub>r</sub> , %	25.2	4.18	26.5	4.37
About 10 <sup>4</sup> CFU/g Lot 2	0 <sup>a</sup>	1	78000	4.89	70500	4.85
		2	80000	4.90	76000	4.88
		3	77000	4.89	78000	4.89
		4	99000	5.00	96000	4.98
		5	55000	4.74	42500	4.63
		Mean	77800	4.88	72600	4.85
		s <sub>r</sub>	15600	0.09	19400	0.13
		RSD <sub>r</sub> , %	20.1	1.87	26.7	2.72
About 10 <sup>6</sup> CFU/g Lot 3	0 <sup>a</sup>	1	6400000	6.81	5850000	6.77
		2	9100000	6.96	8850000	6.95
		3	9350000	6.97	9150000	6.96
		4	6800000	6.83	6400000	6.81
		5	9300000	6.97	8250000	6.92
		Mean	8190000	6.91	7700000	6.88
		s <sub>r</sub>	1460000	0.08	1490000	0.09
		RSD <sub>r</sub> , %	17.8	1.17	19.3	1.27

<sup>a</sup> Naturally contaminated samples.

Table 2. AOAC method comparison (raw pork) (1)

Level	Spike, CFU/g	Replicate	Compact Dry TCaerobic bacteria		AOAC 966.23 aerobic bacteria	
			CFU/g	Log <sub>10</sub> CFU/g	CFU/g	Log <sub>10</sub> CFU/g
About 10 <sup>2</sup> CFU/g Lot 1	0 <sup>a</sup>	1	125	2.10	165	2.22
		2	180	2.26	170	2.23
		3	100	2.00	90	1.95
		4	145	2.16	100	2.00
		5	120	2.08	170	2.23
		Mean	134	2.12	139	2.13
		s <sub>r</sub>	30.3	0.10	40.4	0.14
		RSD <sub>r</sub> , %	22.6	4.52	29.0	6.46
About 10 <sup>4</sup> CFU/g Lot 2	0 <sup>a</sup>	1	37500	4.57	37500	4.57
		2	35000	4.54	35000	4.54
		3	96500	4.98	96500	4.98
		4	16100	4.21	15800	4.20
		5	18350	4.26	10400	4.02
		Mean	40690	4.51	39040	4.46
		s <sub>r</sub>	32600	0.31	34200	0.37
		RSD <sub>r</sub> , %	80.2	6.85	87.6	8.38
About 10 <sup>6</sup> CFU/g Lot 3	0 <sup>a</sup>	1	3700000	6.57	3500000	6.54
		2	2855000	6.46	2840000	6.45
		3	7800000	6.89	7550000	6.88
		4	1210000	6.08	1300000	6.11
		5	1545000	6.19	1775000	6.25
		Mean	3420000	6.44	3393000	6.45
		s <sub>r</sub>	2640000	0.32	2480000	0.29
		RSD <sub>r</sub> , %	77.3	4.98	73.1	4.56

<sup>a</sup> Naturally contaminated samples.

**Table 3. AOAC method comparison (raw lamb) (1)**

Level	Spike, CFU/g	Replicate	Compact Dry TCAerobic bacteria		AOAC 966.23 aerobic bacteria	
			CFU/g	Log <sub>10</sub> CFU/g	CFU/g	Log <sub>10</sub> CFU/g
About 10 <sup>2</sup> CFU/g Lot 1	0 <sup>a</sup>	1	255	2.41	225	2.35
		2	370	2.57	360	2.56
		3	455	2.66	410	2.61
		4	145	2.16	130	2.11
		5	840	2.92	965	2.98
		Mean	413	2.54	418	2.52
		s <sub>r</sub>	266	0.28	325	0.32
		RSD <sub>r</sub> , %	64.4	11.2	77.8	12.8
About 10 <sup>4</sup> CFU/g Lot 2	0 <sup>a</sup>	1	19450	4.29	18900	4.28
		2	13400	4.13	13600	4.13
		3	83500	4.92	79000	4.90
		4	15000	4.18	14200	4.15
		5	10300	4.01	9700	3.99
		Mean	28330	4.31	27100	4.29
		s <sub>r</sub>	31000	0.36	29200	0.36
		RSD <sub>r</sub> , %	109	8.33	108	8.28
About 10 <sup>6</sup> CFU/g Lot 3	0 <sup>a</sup>	1	5850000	6.77	4800000	6.68
		2	1610000	6.21	1120000	6.05
		3	2020000	6.31	2025000	6.31
		4	3700000	6.57	3500000	6.54
		5	1290000	6.11	1320000	6.12
		Mean	2890000	6.39	2553000	6.34
		s <sub>r</sub>	1900000	0.27	1565118.206	0.27
		RSD <sub>r</sub> , %	65.5	4.23	61.30506096	4.26

<sup>a</sup> Naturally contaminated samples.

Table 4. AOAC method comparison (raw veal) (1)

Level	Spike, CFU/g	Replicate	Compact Dry TC aerobic bacteria		AOAC 966.23 aerobic bacteria	
			CFU/g	Log <sub>10</sub> CFU/g	CFU/g	Log <sub>10</sub> CFU/g
0 CFU/g Lot 1	None	1	5	0.70	5	0.70
		2	10	1.00	5	0.70
		3	20	1.30	10	1.00
		4	10	1.00	5	0.70
		5	10	1.00	10	1.00
		Mean	11	1.00	7	0.82
		s <sub>r</sub>	5.48	0.21	2.74	0.16
		RSD <sub>r</sub> , %	49.8	21.3	39.1	20.1
About 10 <sup>2</sup> CFU/g Lot 1	250	1	260	2.41	256	2.41
		2	224	2.35	238	2.38
		3	170	2.23	172	2.24
		4	292	2.47	284	2.45
		5	220	2.34	230	2.36
		Mean	233.2	2.36	236	2.37
		s <sub>r</sub>	45.9	0.09	41.4	0.08
		RSD <sub>r</sub> , %	19.7	3.75	17.5	3.44
About 10 <sup>4</sup> CFU/g Lot 2	25000	1	23200	4.37	23200	4.37
		2	11800	4.07	11900	4.08
		3	13700	4.14	13580	4.13
		4	56000	4.75	53800	4.73
		5	14680	4.17	14800	4.17
		Mean	23900	4.30	23500	4.29
		s <sub>r</sub>	18500	0.27	17500	0.27
		RSD <sub>r</sub> , %	77.4	6.39	74.7	6.21
About 10 <sup>6</sup> CFU/g Lot 3	1500000	1	1340000	6.13	1320000	6.12
		2	1780000	6.25	1540000	6.19
		3	860000	5.93	1210000	6.08
		4	1100000	6.04	1290000	6.11
		5	1540000	6.19	1340000	6.13
		Mean	1324000	6.11	1340000	6.13
		s <sub>r</sub>	361000	0.12	122000	0.04
		RSD <sub>r</sub> , %	27.2	2.03	9.12	0.63

Table 6. Method comparison (independent laboratory data) (1)

Level	Spike, CFU/g	Replicate	Compact Dry TC aerobic bacteria		AOAC 966.23 aerobic bacteria	
			CFU/g	Log <sub>10</sub> CFU/g	CFU/g	Log <sub>10</sub> CFU/g
About 10 <sup>4</sup> CFU/g Lot 1	0 <sup>a</sup>	1	15300	4.18	25400	4.40
		2	9600	3.98	24650	4.39
		3	10900	4.04	20200	4.31
		4	11100	4.05	27950	4.45
		5	10600	4.03	20700	4.32
		Mean	11500	4.06	23780	4.37
		s <sub>r</sub>	2200	0.08	3280	0.06
		RSD <sub>r</sub> , %	19.1	1.89	13.8	1.38
About 10 <sup>6</sup> CFU/g Lot 2	0 <sup>a</sup>	1	1590000	6.20	1300000	6.11
		2	1660000	6.22	1335000	6.13
		3	1870000	6.27	1780000	6.25
		4	1110000	6.05	980000	5.99
		5	1850000	6.27	1680000	6.23
		Mean	1616000	6.20	1415000	6.14
		s <sub>r</sub>	307000	0.09	321000	0.10
		RSD <sub>r</sub> , %	19.0	1.49	22.7	1.68
About 10 <sup>8</sup> CFU/g Lot 3	0 <sup>a</sup>	1	93000000	7.97	86500000	7.94
		2	125000000	8.10	111500000	8.05
		3	164000000	8.21	128000000	8.11
		4	120000000	8.08	104000000	8.02
		5	190000000	8.28	158500000	8.20
		Mean	138400000	8.13	117700000	8.06
		s <sub>r</sub>	38400000	0.12	27250000	0.10
		RSD <sub>r</sub> , %	27.7	1.50	23.2	1.22

<sup>a</sup> Naturally contaminated samples.

#### DISCUSSION OF MODIFICATION STUDY APPROVED APRIL 2015 (4)

For this matrix extension study, the Compact Dry TC was evaluated at two time points, 48 ± 3 h, which is recommended by Nissui and 72 ± 3 h, which is consistent with ISO 4833. Results from each time point were compared to ISO 4833 and to each other. Statistical analyses indicate that there were no significant differences in enumeration of total aerobic counts in cooked chicken, lettuce, frozen fish, milk powder and pasteurized milk between the two time points tested, and thus 48 h is still recommended.

In the single laboratory matrix studies for cooked chicken, lettuce, frozen fish, and milk powder, statistical differences were found between the Compact Dry TC and ISO 4833 for two levels each out of five levels tested for lettuce and frozen fish. In both matrixes, the differences were found in the mid and high contamination levels. For the 48 h Compact Dry TC time point, the mean log<sub>10</sub> differences were >0.5 (0.702 and 0.624 for lettuce, and 0.708 and 0.522 for frozen fish), and although the CIs were small [(0.601, 0.802) and (0.529, 0.719) for lettuce, and (0.538, 0.879) and (0.448, 0.595) for fish], the CIs were outside (-0.5, 0.5) acceptance criterion. In each case, the reference method recorded higher results than the Compact Dry TC. For the other three levels tested for lettuce and fish, mean differences were small (<0.5) and CIs were within (-0.5, 0.5). No significant differences were found in cooked chicken and milk powder. Statistical outliers were detected using Grubb's test in cooked chicken and lettuce, but no justifiable cause was noted in the study for removing the outliers, and so all data were included in the analysis. The r<sup>2</sup> was >0.97 for all matrixes.

In the multi-laboratory study on pasteurized milk, no statistical differences were found between any of the method comparisons after Laboratory 4 was removed from the analysis for unexpectedly high counts in the uncontaminated samples. Due to shipping and scheduling issues, four laboratories initiated testing one day later than the other ten laboratories. An initial review of the data indicated no significant differences in results between the two start dates, so all laboratories (with the exception of Laboratory 4) were included in the statistical analysis. Across 14 data sets, the mean differences between Compact Dry TC at 48 and 72 h were -0.001 to 0.001 (r<sup>2</sup> = 1.0), and the mean differences between the Compact Dry TC at 48 h and ISO 4833 were -0.60 to 0.101 (r<sup>2</sup> = 0.98). The CIs on the mean differences were well within the (-0.5, 0.5) acceptance criterion, and even smaller (less than -0.2, 0.2) for the contaminated samples. The s<sub>r</sub>, RSD<sub>r</sub>, s<sub>R</sub>, and RSD<sub>R</sub> were equally small. These data indicate that the Compact Dry TC gives consistent results, comparable to the reference method, across laboratories.

**MODIFICATION DATA APPROVED APRIL 2015 (4)**

**Table 1. Single laboratory matrix study: Compact Dry TC 48 h vs ISO 4833 (4)**

Matrix	Cont. level	Compact Dry TC 48 h			ISO 4833			Mean diff. <sup>d</sup>	95% CI <sup>e</sup>		r <sup>2h</sup>
		Mean <sup>a</sup>	s <sub>r</sub> <sup>b</sup>	RSD <sub>r</sub> <sup>c</sup>	Mean	s <sub>r</sub>	RSD <sub>r</sub>		LCL <sup>f</sup>	UCL <sup>g</sup>	
Cooked chicken	1	0.281	0.594	211	0.357	0.757	212	0.076	-0.070	0.219	0.99
	2	2.399	0.304	12.7	2.572	0.517	20.1	0.173	-0.113	0.459	
	3	3.309	0.364	11.0	3.383	0.357	10.6	0.074	-0.015	0.163	
	4	5.056	0.284	5.62	5.404	0.133	2.46	0.348	0.207	0.490	
	5	6.173	0.252	4.08	6.194	0.209	3.37	0.021	-0.036	0.077	
Lettuce	1	2.346	0.663	28.3	2.663	0.682	25.6	0.317	0.156	0.478	0.98
	2	3.359	0.709	21.1	3.437	0.346	10.1	0.078	-0.230	0.387	
	3	4.491	0.324	7.21	5.193	0.352	6.78	0.702	0.601	0.802	
	4	5.443	0.201	3.69	5.638	0.244	4.28	0.196	0.129	0.263	
	5	7.010	0.143	2.04	7.634	0.134	1.76	0.624	0.529	0.719	
Frozen fish	1	2.953	0.178	6.03	3.089	0.158	5.11	0.136	0.016	0.256	0.97
	2	3.924	0.121	3.08	4.052	0.091	2.25	0.128	0.056	0.200	
	3	4.791	0.191	3.99	5.499	0.139	2.53	0.708	0.538	0.879	
	4	5.854	0.135	2.31	6.246	0.091	1.46	0.392	0.326	0.459	
	5	6.432	0.066	1.03	6.953	0.090	1.29	0.522	0.448	0.595	
Milk powder	1	0.645	0.845	131	1.012	0.736	72.7	0.367	-0.067	0.800	0.98
	2	2.064	0.139	6.73	2.232	0.157	7.03	0.168	0.082	0.253	
	3	4.026	0.134	3.33	3.993	0.076	1.90	-0.033	-0.102	0.036	
	4	5.010	0.110	2.20	4.926	0.103	2.09	-0.084	-0.149	-0.018	
	5	5.709	0.132	2.31	5.639	0.171	3.03	-0.070	-0.127	-0.012	

<sup>a</sup>Mean of five replicate portions, plated in duplicate, after logarithmic transformation: Log<sub>10</sub>[CFU/g + (0.1)f].

<sup>b</sup>Repeatability standard deviation.

<sup>c</sup>Relative standard deviation for repeatability.

<sup>d</sup>Mean difference between the candidate and reference methods.

<sup>e</sup>Confidence interval.

<sup>f</sup>95% Lower confidence limit for difference of means.

<sup>g</sup>95% Upper confidence limit for difference of means.

<sup>h</sup>Square of correlation coefficient.



Table 2. Single laboratory matrix study: Compact Dry TC 72 h vs ISO 4833 (4)

Matrix	Cont. level	Compact Dry TC 72 h			ISO 4833			Mean diff. <sup>d</sup>	95% CI <sup>e</sup>		r <sup>2h</sup>
		Mean <sup>a</sup>	s <sub>r</sub> <sup>b</sup>	RSD <sub>r</sub> <sup>c</sup>	Mean	s <sub>r</sub>	RSD <sub>r</sub>		LCL <sup>f</sup>	UCL <sup>g</sup>	
Cooked chicken	1	0.281	0.594	211	0.357	0.757	212	0.076	-0.068	0.219	0.99
	2	2.734	0.636	23.4	2.572	0.517	20.1	-0.163	-0.358	0.033	
	3	3.323	0.369	11.1	3.383	0.357	10.6	0.060	-0.033	0.153	
	4	5.063	0.280	5.53	5.404	0.133	2.46	0.342	0.203	0.480	
	5	6.179	0.249	4.03	6.194	0.209	3.37	0.015	-0.040	0.070	
Lettuce	1	2.496	0.593	23.8	2.663	0.682	25.6	0.168	0.040	0.296	0.98
	2	3.393	0.732	21.6	3.437	0.346	10.1	0.044	-0.277	0.365	
	3	4.782	0.246	5.14	5.193	0.352	6.78	0.411	0.309	0.513	
	4	5.527	0.200	3.62	5.638	0.244	4.28	0.111	0.034	0.188	
	5	7.117	0.182	2.56	7.634	0.134	1.76	0.517	0.419	0.615	
Frozen fish	1	3.000	0.155	5.17	3.089	0.158	5.11	0.090	-0.019	0.198	0.98
	2	3.924	0.121	3.08	4.052	0.091	2.25	0.128	0.056	0.199	
	3	4.822	0.174	3.61	5.499	0.139	2.53	0.677	0.518	0.838	
	4	5.860	0.133	2.27	6.246	0.091	1.46	0.386	0.323	0.450	
	5	6.474	0.065	1.00	6.953	0.090	1.29	0.479	0.414	0.544	
Milk powder	1	0.881	0.792	89.9	1.012	0.736	72.7	0.130	-0.190	0.450	0.99
	2	2.105	0.154	7.32	2.232	0.157	7.03	0.127	0.026	0.277	
	3	4.027	0.133	3.30	3.993	0.076	1.90	-0.034	-0.103	0.035	
	4	5.010	0.110	2.20	4.926	0.103	2.09	-0.084	-0.149	-0.018	
	5	5.720	0.131	2.29	5.639	0.171	3.03	-0.081	-0.142	-0.019	

<sup>a</sup>Mean of five replicate portions, plated in duplicate, after logarithmic transformation:  $\text{Log}_{10}[\text{CFU/g} + (0.1)\text{f}]$ .

<sup>b</sup>Repeatability standard deviation.

<sup>c</sup>Relative standard deviation for repeatability.

<sup>d</sup>Mean difference between the candidate and reference methods.

<sup>e</sup>Confidence interval.

<sup>f</sup>95% Lower confidence limit for difference of means.

<sup>g</sup>95% Upper confidence limit for difference of means.

<sup>h</sup>Square of correlation coefficient.

**Table 3. Single laboratory matrix study: Compact Dry TC 48 h vs Compact Dry 72 h (4)**

Matrix	Cont. level	Compact Dry TC 48 h			Compact Dry TC 72 h			Mean diff. <sup>d</sup>	95% CI <sup>e</sup>		r <sup>2h</sup>
		Mean <sup>a</sup>	s <sub>r</sub> <sup>b</sup>	RSD <sub>r</sub> <sup>c</sup>	Mean	s <sub>r</sub>	RSD <sub>r</sub>		LCL <sup>f</sup>	UCL <sup>g</sup>	
Cooked chicken	1	0.281	0.594	211	0.281	0.594	211	0.00	0.00	0.00	0.99
	2	2.399	0.304	12.7	2.734	0.636	23.4	0.336	-0.079	0.750	
	3	3.309	0.364	11.0	3.323	0.369	11.1	0.014	-0.001	0.029	
	4	5.056	0.284	5.62	5.063	0.280	5.53	0.007	0.002	0.011	
	5	6.173	0.252	4.08	6.179	0.249	4.03	0.006	0.003	0.009	
Lettuce	1	2.346	0.663	28.3	2.496	0.593	23.76	0.149	0.065	0.233	0.99
	2	3.359	0.709	21.1	3.393	0.732	21.6	0.034	0.007	0.061	
	3	4.491	0.324	7.21	4.782	0.246	5.14	0.291	0.200	0.381	
	4	5.443	0.201	3.69	5.527	0.200	3.62	0.085	0.001	0.169	
	5	7.010	0.143	2.04	7.117	0.182	2.56	0.107	0.031	0.183	
Frozen fish	1	2.953	0.178	6.03	3.000	0.155	5.17	0.046	0.026	0.067	1.00
	2	3.924	0.121	3.08	3.924	0.121	3.08	0.000	-0.001	0.000	
	3	4.791	0.191	3.99	4.822	0.174	3.61	0.031	0.000	0.063	
	4	5.854	0.135	2.31	5.860	0.133	2.27	0.006	0.002	0.010	
	5	6.432	0.066	1.03	6.474	0.065	1.00	0.043	-0.022	0.107	
Milk powder	1	0.645	0.845	131	0.881	0.792	89.9	0.236	-0.123	0.596	0.99
	2	2.064	0.139	6.73	2.105	0.154	7.32	0.041	0.008	0.074	
	3	4.026	0.134	3.33	4.027	0.133	3.30	0.001	0.000	0.003	
	4	5.010	0.110	2.20	5.010	0.110	2.20	0.000	0.000	0.001	
	5	5.709	0.132	2.31	5.720	0.131	2.29	0.011	0.002	0.020	

<sup>a</sup>Mean of five replicate portions, plated in duplicate, after logarithmic transformation: Log<sub>10</sub>[CFU/g + (0.1)f].

<sup>b</sup>Repeatability standard deviation.

<sup>c</sup>Relative standard deviation for repeatability.

<sup>d</sup>Mean difference between the candidate and reference methods.

<sup>e</sup>Confidence interval.

<sup>f</sup>95% Lower confidence limit for difference of means.

<sup>g</sup>95% Upper confidence limit for difference of means.

<sup>h</sup>Square of correlation coefficient.

**Table 4. Multi-laboratory study: Summary of pasteurized milk results by laboratory for each contamination level for each method (4)**

Laboratory	Uncontaminated			Low level			Medium level			High level		
	CD TC <sup>a</sup> 48 h	CD TC 72 h	ISO 4833	CD TC 48 h	CD TC 72 h	ISO 4833	CD TC 48 h	CD TC 72 h	ISO 4833	CD TC 48 h	CD TC 72 h	ISO 4833
1	0.000 <sup>b</sup>	0.000	0.734	2.557	2.560	2.698	3.533	3.538	3.697	4.701	4.703	4.723
2	0.000	0.000	0.000	2.638	2.638	2.832	3.726	3.726	3.825	4.727	4.727	4.860
3	0.521	0.591	0.000	2.684	2.689	2.806	3.763	3.771	3.973	4.803	4.808	4.947
4	2.340	2.365	2.192	2.930	2.937	2.992	3.872	3.909	3.943	4.829	4.852	5.049
5	0.000	0.000	0.000	2.599	2.599	2.762	3.728	3.740	3.805	4.740	4.748	4.814
6	0.000	0.000	0.000	2.795	2.810	2.694	3.737	3.766	3.758	4.811	4.816	4.832
7	0.000	0.000	0.000	2.749	2.749	2.852	3.831	3.831	3.953	4.823	4.825	4.945
8	0.000	0.000	0.000	2.588	2.588	2.770	3.627	3.627	3.917	4.761	4.761	4.826
9	0.260	0.260	0.000	2.751	2.751	2.803	3.790	3.790	3.888	4.904	4.904	4.963
10	0.000	0.260	0.000	2.618	2.618	2.788	3.688	3.688	3.724	4.710	4.710	4.616
11 <sup>c</sup>	0.000	0.000	0.000	2.703	2.705	2.752	3.726	3.726	3.820	4.793	4.793	4.833
12 <sup>c</sup>	0.000	0.000	0.000	2.748	2.753	2.730	3.812	3.734	3.854	4.983	4.940	4.979
13 <sup>c</sup>	0.000	0.000	0.000	2.779	2.779	2.825	3.876	3.878	3.896	4.853	4.858	4.864
14 <sup>d</sup>	0.260	0.260	0.000	2.700	2.700	2.810	3.739	3.739	3.789	4.835	4.836	4.833
14 <sup>c,d</sup>	0.260	0.260	0.000	2.700	2.700	2.745	3.774	3.774	3.833	4.716	4.716	4.701

<sup>a</sup>Compact Dry TC.

<sup>b</sup>Results are reported for each laboratory as a mean of 2 replicate portions, plated in duplicate, after logarithmic transformation:  $\text{Log}_{10}[\text{CFU/g} + (0.1)\text{f}]$ .

<sup>c</sup>Sample analysis initiated one day later than scheduled start date.

<sup>d</sup>Organizing Laboratory.

**Table 6. Multi-laboratory study: Method comparison of pasteurized milk results for each contamination level across 14 data sets<sup>a</sup> (4)**

Contam. level	Compact Dry TC 48 h vs ISO 4833						Compact Dry TC 72 h vs ISO 4833						Compact Dry TC 48 h vs 72 h					
	Mean <sup>b</sup> CD TC 48 h	Mean ISO 4833	Mean diff. <sup>c</sup>	LCL <sup>d</sup>	UCL <sup>e</sup>	r <sup>2f</sup>	Mean CD TC 72 h	Mean ISO 4833	Mean diff.	LCL	UCL	r <sup>2</sup>	Mean CD TC 48 h	Mean CD TC 72 h	Mean diff.	LCL	UCL	r <sup>2</sup>
Uncont.	0.112	0.052	-0.060	-0.270	0.080		0.117	0.052	-0.065	-0.281	0.079		0.112	0.117	0.005	-0.006	0.018	
Low	2.686	2.776	0.090	0.046	0.129	0.98	2.688	2.776	0.088	0.043	0.129	0.98	2.686	2.688	0.002	-0.003	0.006	1.0
Mid	3.737	3.838	0.101	0.064	0.131		3.738	3.838	0.100	0.065	0.134		3.737	3.738	0.001	-0.011	0.008	
High	4.797	4.838	0.041	0.011	0.069		4.796	4.838	0.042	0.012	0.070		4.797	4.796	-0.001	-0.008	0.006	

<sup>a</sup>Laboratory 4 removed from statistical analysis due to unexpectedly high counts in the uncontaminated samples. See Table 4.

<sup>b</sup>Overall mean across laboratories for each method, after logarithmic transformation:  $\text{Log}_{10}[\text{CFU}/\text{g} + (0.1)\text{f}]$ .

<sup>c</sup>Mean difference between methods.

<sup>d</sup>95% Lower confidence limit for difference of means.

<sup>e</sup>95% Upper confidence limit for difference of means.

<sup>f</sup>Square of correlation coefficient.

#### REFERENCES CITED

1. Kodaka, H., Mizuochi, S., Teramura, H., and Nirazuka, T., Comparison of the Compact Dry TC Method with the Standard Pour Plate Method (AOAC Official Method 966.23) for Determining Aerobic Colony Counts in Food Samples, AOAC® Performance Tested<sup>SM</sup> certification number 010404.
2. *Official Methods of Analysis* 19<sup>th</sup> Ed., AOAC INTERNATIONAL, Rockville, MD, Method **966.23**, *Microbiological Methods*, <http://www.eoma.aoac.org>, accessed 2004
3. ISO 4833:2003, *Microbiology of food and animal feeding stuffs – Horizontal method for the enumeration of microorganisms – Colony-count technique at 30°C*, [http://www.iso.org/iso/iso\\_catalogue/catalogue\\_tc/catalogue\\_detail.htm?csnumber=34524](http://www.iso.org/iso/iso_catalogue/catalogue_tc/catalogue_detail.htm?csnumber=34524), accessed 2006
4. Mizuochi, S., Evaluation of Compact Dry TC: Matrix Extension, AOAC® Performance Tested<sup>SM</sup> certification number 010404. Modification approved April 2015.