



## CERTIFICATE OF COMPLIANCE

### LRQA

hereby declares that the certification assessment has demonstrated that

### COMPACT DRY X-SA

Manufactured and supplied by:  
Shimadzu Diagnostics Corporation  
20th Floor Ueno Frontier Tower  
3-24-6 Ueno, Taito-ku, Tokyo, 110-8736  
JAPAN

has been validated and revealed to be at least equivalent to the reference method as demonstrated by the validation study report. The summary of the validation report is available on the MicroVal website: [www.microval.org](http://www.microval.org)

#### Reference method:

EN ISO 6888-1:2021 Microbiology of the food chain — Horizontal method for the enumeration of coagulase-positive staphylococci (*Staphylococcus aureus* and other species) — Part 1: Technique using Baird-Parker agar medium

Scope: A broad range of foods, pet food and animal feed and environmental samples

The validation and certification has been performed in accordance with EN ISO 16140-2:2016 and the MicroVal Rules and Certification Scheme version 9.1.

Certificate no.: 2008-LR14

First approval date: 25 May 2011  
Surveillance date: 21 September 2023  
Expiry date: 24 May 2027

ISSUED BY: \_\_\_\_\_  
LRQA Nederland B.V.  
Rotterdam, The Netherlands